

## 1. Overview of Pulsed Light Measurement

#### Why use pulsed light?

A semiconductor laser emits light by supplying a driving current. However, the drive current causes the semiconductor laser chip itself to generate heat, causing a shift in the output wavelength. In the worst case, the heat can damage the laser. To avoid such problems, pulse-driven light emission is widely used. In addition, in the case of lasers for material processing that require a high-power laser beam, in order to generate a high-power laser beam, the power is accumulated for a certain period of time, and then it is emitted with a short pulse. The pulse drive conditions of these lasers are designed with repetition frequency and duty ratio depending on the application. In order to accurately measure these pulsed lights using an optical spectrum analyzer (OSA), it is necessary to understand the characteristics of the OSA and select the appropriate measurement method and settings.

#### Challenges of pulsed light measurement by OSA

OSA samples and measures the total power of a specific wavelength segment while sweeping the wavelength, and displays it as an optical spectrum. Normally, it is assumed that the optical input signal is constant during the OSA wavelength sweep. In other words, a signal that turns the optical input signal on / off like pulsed light is not expected. Therefore, if the OSA measurement method and condition settings are not suitable for the pulsed light drive conditions, the measured spectrum may appear segmented or choppy as shown below.

### Pulsed input



OSA trace



### Pulsed light measurement by Yokogawa OSA

Yokogawa's OSA offers three types of measurement methods, "time average spectrum measurement", "peak hold measurement", and "external trigger synchronous measurement", depending on the driving conditions of the pulsed light.

#### • Time average spectrum measurement

The average power of the pulsed light is measured as the power of the light spectrum. It can be used even when the pulse width is narrow.

#### • External trigger synchronous measurement

By supplying the pulsed signal from the light source as an external electrical trigger signal synchronized with the optical pulse to the OSA, the peak value of the optical pulse is captured. The measurable pulse width is 50  $\mu$ s or more.

Peak hold measurement

Captures the peak value within the repeating cycle of the optical pulse. The measurable pulse width is 100  $\mu s$  or more.

#### Applicable measurement modes per OSA model

Measurement modes	AQ6360	AQ6370 AQ6373 AQ6374	AQ6375 AQ6376 AQ6377	
Time average spectrum	٠	•	•	
External trigger		•	0	•: Available
Peak hold		•	0	Blank: Not available

Note. The external trigger and peak hold measurements are not available in HIGH1, HIGH2, and HIGH3 of AQ6375, AQ6376, and AQ6377 where the high dynamic range "CHOP" mode is used.



### 2. Pulsed Light Measurement in Time Average Mode (normal mode)

In the time average mode, the time average power of pulsed light is measured as the optical spectral power of each wavelength. When the pulsed light is a square wave, the measured average light power is calculated as (peak power of light pulse [mW])  $\times$  (duty ratio of pulsed light). Therefore, the smaller the duty ratio of pulsed light, the lower the measured power. In order to measure the average power correctly in this mode, the repetition frequency must be high to some extent. Otherwise, it may need to be measured in a high sensitivity setting or by increasing the number of averages.

#### Features of time average mode

- No external trigger signal required.
- Cannot handle irregular pulse signals. (Because the average power fluctuates)
- It can handle any signal such as short pulse and long repetition period.
- High power pulsed light with peak power exceeding 1W can be measured. However, there are limits on maximum pulse peak power and maximum safe average power.
- The measured optical power is low, because the average power of the optical pulse is measured. Therefore, there are likely to be cases where measurement is not possible due to insufficient power. To ensure a high dynamic range, it is necessary to use a high sensitivity setting, which slows down the measurement. If the average power is high, it could measure faster than other methods.
- There are restrictions on sensitivity setting depending on the repetition frequency. Therefore, when the repetition frequency is low, the measurement sensitivity needs to be increased, which slows down the measurement.
- High dynamic range mode (CHOP), double speed mode, averaging function can be applied.

#### Constraints on sensitivity setting by repetition frequency

The pulse repetition frequency can be roughly divided into the following three groups, and the measurement method can be considered accordingly.

#### • Pulse repetition frequency: MHz

At very high pulse repetition frequencies, the reaction speed of the detection circuit in the OSA is slow enough even at the lowest sensitivity setting (NORM HOLD), so the OSA can measure it like a continuous wave regardless of the device setting. Therefore, set the sensitivity according to the average power.

#### • Pulse repetition frequency: kHz

For kHz class pulse repetition frequencies, on / off distortion due to pulsed light may be noticeable. This effect can be eliminated by selecting a high sensitivity setting and slowing down the reaction speed of the detection circuit. Therefore, it is necessary to set the sensitivity considering both the average power and the repetition frequency.

#### • Pulse repetition frequency: Hz

If the pulse repetition frequency is very low, set the sensitivity considering the repetition frequency. If the highest sensitivity setting does not eliminate the distortion of the recorded spectrum, add an averaging process. Assuming that the number of averaging is n, the minimum repetition frequency of each sensitivity setting is multiplied by 1 / n. However, adding the number of averaging will significantly reduce the measurement speed, so use it only when the maximum sensitivity setting is insufficient.

#### Estimated minimum repetition frequency

The table on the right shows the minimum repeat frequency that can be measured with each sensitivity setting.

Note. HIGH1-3 of AQ6375/76/77 is in high dynamic range mode (CHOP), and the averaging does not reduce the minimum repeat frequency in this mode.

	Min.	repetition	freq.
Sensitivity setting	AQ6360	AQ6370 AQ6373 AQ6374	AQ6375 AQ6376 AQ6377
NORM/HOLD	1 MHz	200 kHz	200 kHz
NORM/AUTO	1 MHz	100 kHz	100 kHz
NORMAL	200 kHz	33 kHz	33 kHz
MID	50 kHz	10 kHz	10 kHz
HIGH1	10 kHz	3.3 kHz	1 MHz
HIGH2	2 kHz	660 Hz	1 MHz
HIGH3	N/A	160 Hz	1 MHz

#### Optical input power conditions

The time average mode can measure high power pulsed light with a peak power of more than 1W. However, pulsed light of which average spectral power exceeds the OSA maximum input power specification cannot be measured.

\* Maximum input power is the maximum spectral power per measurement resolution.

Max, input	AQ6360	0 AQ6370	AQ6373		AQ6374			100070	
power			400 – 550	550 – 1100	400 – 550	550 – 1750 nm	AQ6375	AQ6376	AQ6377
dBm	+20	+20	+10	+20	+10	+20	+20	+13	+13
W	0.1	0.1	0.01	0.1	0.01	0.1	0.1	0.02	0.02

**[Caution]** There is a limit to the optical power that can be input to OSA. See "Guideline for Pulsed Light Input Power" in section 5 for more information.



## 3. Pulsed Light Measurement in External Trigger Mode

The External trigger mode is a mode that uses an external trigger signal to control the timing of OSA measurements. By supplying an external trigger signal synchronized with the optical pulse signal to the OSA, the peak value of the optical pulse can be captured. OSA takes one sample from one laser pulse. OSA increments to the next wavelength step and records the output signal of the detection circuit each time one trigger is input. (See manual for electrical properties of trigger signal)

#### Features of external trigger mode

- Requires an external trigger signal synchronized with the optical pulse.
- It can handle irregular pulse signals.
- The max. optical pulse peak power is the maximum input power of OSA. (For AQ6370, +20 dBm)
- The minimum optical pulse width is 50 µs (for Norm/Hold sensitivity) However, the minimum pulse width depends on the sensitivity setting. Higher sensitivity settings increase the minimum pulse width.
- Since the peak power of the optical pulse is captured, the optical power measured is higher than that of the time average mode. Therefore, problems due to insufficient power are unlikely to occur. In addition, measurement is fast because a high dynamic range can be secured even with a low sensitivity setting.
- There are no restrictions on sensitivity setting depending on the repetition frequency. Therefore, it is easy to speed up the measurement using the low sensitivity setting.
- High dynamic range mode (CHOP), double speed mode, and averaging function cannot be applied.

#### Minimum pulse width condition

The external trigger mode requires a pulse width of at least 50 us to capture the peak of the optical pulse. In addition, the minimum pulse width depends on the measurement sensitivity setting used, and the higher the measurement sensitivity, the wider the minimum pulse width at which peaks are captured. Therefore, the suitability of this mode is mainly determined by the pulse width to be measured. The reason is that the gain of the detection circuit and the response speed are inversely proportional. At low sensitivity settings, the detection circuit gain is low and the response is fast, allowing short pulse width peaks to be measured correctly. However, at high sensitivity settings, the gain of the detection circuit is high and the response is slow, so short pulse width peaks cannot be measured correctly.

#### Estimated minimum pulse width

The table on the right shows the min. pulse width that can be measured with each sensitivity setting.

Note. HIGH1-3 of AQ6375/76/77 is in high dynamic range mode (CHOP).

The AQ6360 does not have the external trigger mode.

#### External trigger and timing adjustment

	Min. pu	Ise width
Sensitivity	AQ6370	AQ6375
setting	AQ6373	AQ6376
	AQ6374	AQ6377
NORM/HOLD	50 µs	50 µs
NORM/AUTO	300 µs	300 µs
NORMAL	1 ms	1 ms
MID	3 ms	3 ms
HIGH1	10 ms	-
HIGH2	50 ms	-
HIGH3	200 ms	-

To measure the peak power of the optical pulse, connect an external trigger signal synchronized with the optical pulse to the [TRIGGER IN] port on the back of the OSA, and set [TRIG INPUT MODE] in the OSA SYSTEM menu to [SMPL TRIG MODE]. The SMPL TRIGGER MODE triggers at the rising edge (or falling edge) of the input external trigger signal to measure 1 point. Adjust the timing as needed to capture the peak of the optical pulse correctly. OSA has a delay of about 70 µs from trigger to measurement. It also has a function to add a delay time in the range of 0 to 1000 µs.

#### Optical input power conditions

In external trigger mode, OSA captures the peak power of the optical pulse, so optical pulsed light that exceeds the OSA maximum input power specifications cannot be measured. Therefore, the peak power of a measurable optical pulse is at most +20 dBm (0.1 W). The maximum input power of each model is shown in the table below.

-					_	
Max. input	1002200	400270	AQ6373	AQ6374	A00275	100
	AUDSDU	AU63/0			AU03/3	AUD

\* Maximum input power is the maximum spectral power per measurement resolution.

Max. input	AQ6360	AQ6370	AQ6373		AQ6374		A00275	A06376	100377
power			400 – 550 nm	550 – 1100 nm	400 – 550 nm	550 – 1750 nm	AQ6375	AQ0370	AQ0311
dBm	+20	+20	+10	+20	+10	+20	+20	+13	+13
W	0.1	0.1	0.01	0.1	0.01	0.1	0.1	0.02	0.02

[Caution] There is a limit to the optical power that can be input to OSA. See "Guideline for Pulsed Light Input Power" in section 5 for more information.



## 4. Pulsed Light Measurement in Peak Hold Mode

Peak hold mode is a measurement mode that does not require an external trigger signal. In peak hold mode, the detection signal is recorded for the specified period for each measurement wavelength, and the maximum value of the data acquired during that period is used as the power for that measurement wavelength. The period during which the detection signal is recorded is called the "hold time" and is set to a value larger than the pulse repetition period (1 / repetition frequency). Which allows to measure at least one pulse within the hold time. OSA basically takes one sample from one laser pulse and repeats the movement of the measurement wavelength and the recording of the output signal of the detection circuit.

#### Features of peak hold mode

- No external trigger signal required.
- The max. optical pulse peak power is the maximum input power of OSA. (For AQ6370, +20 dBm)
- The minimum optical pulse width is 100 µs (for Norm/Hold) However, the minimum pulse width depends on the sensitivity setting. Higher sensitivity settings increase the minimum pulse width.
- Since the peak power of the optical pulse is captured, the optical power measured is higher than that of the time average mode. Therefore, problems due to insufficient power are unlikely to occur. In addition, measurement is fast because a high dynamic range can be secured even with a low sensitivity setting.
- There are no restrictions on sensitivity setting depending on the repetition frequency. Therefore, it is easy to speed up the measurement using the low sensitivity setting.
- High dynamic range mode (CHOP), double speed mode, and averaging function cannot be applied.
- Not recommended for measuring irregular pulse signals. Because it is necessary to set the hold time according to the minimum repetition frequency.

#### Minimum pulse width condition

The peak hold mode requires a pulse width of at least 100  $\mu$ s to capture the peak of the optical pulse. In addition, the minimum pulse width depends on the measurement sensitivity setting used, and the higher the measurement sensitivity, the wider the minimum pulse width at which peaks are captured. Therefore, the suitability of this mode is mainly determined by the pulse width to be measured. The reason is that the gain of the detection circuit and the response speed are inversely proportional.

At low sensitivity settings, the detection circuit gain is low and the response is fast, allowing short pulse width peaks to be measured correctly. However, at high sensitivity settings, the gain of the detection circuit is high and the response is slow, so short pulse width peaks cannot be measured correctly.

#### Estimated minimum pulse width

The table on the right shows the min. pulse width that can be measured with each sensitivity setting.

Note. HIGH1-3 of AQ6375/76/77 is in high dynamic range mode (CHOP).

The AQ6360 does not have the peak hold mode.

	Min. pu	i. pulse width			
Sensitivity	AQ6370	AQ6375			
setting	AQ6373	AQ6376			
	AQ6374	AQ6377			
NORM/HOLD	100 µs	100 µs			
NORM/AUTO	300 µs	300 µs			
NORMAL	1 ms	1 ms			
MID	3 ms	3 ms			
HIGH1	10 ms	-			
HIGH2	50 ms	-			
HIGH3	200 ms	-			

#### Hold time setting

Set the hold time to a value larger than the pulse repetition period (1 / repetition frequency). In OSA, the hold time can be set arbitrarily within the range of 0 to 9999 ms. Therefore, it can measure down to a pulse repetition frequency of 0.1 Hz.

#### Optical input power conditions

In Peak hold mode, OSA captures the peak power of the optical pulse, so optical pulsed light that exceeds the OSA maximum input power specifications cannot be measured. Therefore, the peak power of a measurable optical pulse is at most +20 dBm (0.1 W). The maximum input power of each model is shown in the table below.

\* Maximum input power is the maximum spectral power per measurement resolution.

Max. input power	AQ6360	5360 AQ6370	AQ6373		AQ6374		400075	A06376	A06277
			400 – 550 nm	550 – 1100 nm	400 – 550 nm	550 – 1750 nm	AQ6375	AQ6376	AQ6377
dBm	+20	+20	+10	+20	+10	+20	+20	+13	+13
W	0.1	0.1	0.01	0.1	0.01	0.1	0.1	0.02	0.02

**[Caution]** There is a limit to the optical power that can be input to OSA. See "Guideline for Pulsed Light Input Power" in section 5 for more information.



## 5. Guideline for Pulsed Light Input Power

When measuring pulsed light, it is recommended to use the pulse peak power below the maximum safe input power of the OSA to prevent damage inside the equipment. However, since it is considered that the limitation differs depending on the emission condition of the pulsed light, some conditions are described below. It should be noted that these conditions are merely estimated values and do not guarantee safety.

#### When the pulse width exceeds 1 µs

The pulse peak power must not exceed the OSA maximum safe input power.

#### When the pulse width is 1 µs or less

The pulse peak power must not exceed 316W.

The averaged power must not exceed the OSA maximum safe input power.

## The average power is calculated from:

Pulse peak power (W)  $\times$  pulse width (s)  $\div$  repetition cycle (s), or Pulse peak power (W)  $\times$  pulse width (s)  $\times$  repetition frequency (Hz)

#### OSA maximum safe input power

Max. safe input power	AQ6360		400070	AQ	6373	AQ	6374	A00075	100070	100077
		AQ6370	400 – 550 nm	550 – 1100 nm	400 – 550 nm	550 – 1750 nm	AQ6375	AQ6376	AQ6377	
dBm	+25	+25	+10	+20	+10	+20	+25	+20	+20	
W	0.316	0.316	0.01	0.1	0.01	0.1	0.316	0.1	0.1	

\* The above guidelines do not indicate the optical power that can be measured by the OSA.

■ How to obtain the max. pulse peak power from pulse conditions





## 6. Summary of the Guideline

Comply with "Guideline for pulsed light input power."

#### Use the time average mode

Time averaging mode is the easiest and most flexible way to measure pulsed light.

- Consider External trigger mode and Peak hold mode when you encounter issues with Time average mode For example;
  - For irregular pulsed light
  - Measurement is not possible. Because the average power is below the measurement sensitivity due to the relationship between the peak power and the duty ratio.
  - Measurement takes too long because the high sensitivity setting is required due to very low repetition frequency.

### External trigger mode or Peak hold mode

- Determine suitability by the pulse width and the required sensitivity setting.
- Chose one by the availability of an external trigger and the minimum pulse width.

	Time averaging mode ( normal mode)	External trigger mode	Peak hold mode			
External trigger	No	Required	No			
Irregular pulse	Not applicable	Applicable	Applicable but not recommended			
Minimum pulse	No restriction	50 µs (NORM/HOLD)	100 µs (NORM/HOLD)			
width		* It changes depending on t	he sensitivity setting.			
Maximum input power	High power pulse light with peak power exceeding 1W can be measured. However, there are restrictions on the maximum pulse peak power and maximum safe average power. * Refer to the Guidelines for Pulsed Light Input Power. (page 5)	<ul> <li>Pulse peak power must be less than or equal to OSA maximum input power (For AQ6370, 0.1W or less)</li> <li>* Refer to the maximum input power of each product. (page 3 &amp;4)</li> </ul>				
Measured power	Average power	Pulse peak power				
How to select sensitivity	<ul><li>Step 1: By a repetition frequency</li><li>Step 2: By a required dynamic range</li></ul>	<ul><li>Step 1: By a pulse width</li><li>Step 2: By a required dynamic range</li></ul>				
Selection guide line	<ul> <li>Regular pulse signal with a constant repetition frequency</li> <li>Average power is above measurement sensitivity</li> </ul>	<ul> <li>The pulse width satisfies the condition for each sensitivity.</li> <li>* Refer to the guideline for the minimum pulse width for each sensitivity setting. (page 3&amp; 4)</li> </ul>				
		• There is an external trigger signal	-			
Tips	<ul> <li>The measurement time can be shortened by increasing the repetition frequency.</li> <li>When the repetition frequency is extremely low, measurement becomes possible by increasing the number of averaging. However, averaging leads to a significant increase in measurement time, so it is not recommend to use it anything other than the sensitivity setting HIGH3.</li> </ul>	If the repetition frequency faster than the time average	is less than 20 Hz, it may be ge mode.			



## 7. Examples (AQ6370)

#### ■ Pulse width 100 µs, Pulse peak power 0.1 W, Rep-frequency 1kHz

- Any measurement mode can be selected according to the pulse width condition. If there is an external trigger signal, select "external trigger". Otherwise, select "Peak Hold".
- For External trigger and Peak hold, make sure that the pulse peak is below "maximum input power".

Mode	Pulse width	Pul p	se peak oower	Rep freq.	Sensitivity setting	Measured power	Noise level	Dynamic range	
Time Avg					High2	+10 dBm	-85 dBm	95 dB	
Ex-Trigger	100 us	0.1 W	+20 dBm	1 kHz	Norm/Hold	+20 dBm	-25 dBm	45 dB	
Peak Hold						Norm/Hold	+20 dBm	-25 dBm	45 dB

### ■ Pulse width 50 µs, Pulse peak power 0.1 W, Rep-frequency 1kHz

- "Time average" or "external trigger" can be selected according to the pulse width condition. Select "external if there is an external trigger signal since "external trigger" is faster.
- In the case of "external trigger", make sure that the pulse peak is below the maximum input power.
- In the case of "time average", make sure that the pulse peak is below the maximum safe input power according to the pulse width condition, and set the sensitivity setting to High 2 or higher according to the repetition frequency.

Mode	Pulse width	Pu	lse peak power	Rep freq.	Sensitivity setting	Measured power	Noise level	Dynamic range
Time Avg	F0	0 us 0.1 W +20 dBm 1 kH			High2	+7 dBm	-85 dBm	92 dB
Ex-Trigger	50 us			Norm/Hold	+20 dBm	-25 dBm	45 dB	

### ■ Pulse width 1 µs, Pulse peak power 0.1 W, Rep-frequency 1kHz

- Select "Time average" according to the pulse width condition.
- Make sure that the pulse peak is below the maximum safe input power according to the pulse width condition.
- Set the sensitivity setting to High 2 or higher according to the repetition frequency.

Mode	Pulse width	Pulse peak power		Rep freq.	Sensitivity setting	Measured power	Noise level	Dynamic range
Time average	1 µs	0.1 W	+20 dBm	100 kHz	Norm/Hold	+10 dBm	-55 dBm	45 dB
				10 kHz	MID	0 dBm	-75 dBm	75 dB
				1 kHz	High2	-10 dBm	-85 dBm	75 dB
				200 Hz	High3	-17 dBm	-90 dBm	73 dB

#### Pulse width 1 ns, Pulse peak power 10 W, Rep-frequency 1kHz

- Select "Time average" according to the pulse width condition.
- Make sure that the pulse peak is 316 W or less and the average power is below the maximum safe input power according to the pulse width condition.
- Pulse Pulse peak Rep Sensitivity Measured Dynamic Mode Noise level width power freq. setting power range 100 kHz Norm/Hold -65 dBm 0 dBm 45 dB 10 kHz -10 dBm MID -75 dBm 65 dB Time 10 W +40 dBm 1 ns average 1 kHz High2 -20 dBm -85 dBm 65 dB 200 Hz High3 -27 dBm -90 dBm 63 dB
- Set the sensitivity setting to High 2 or higher according to the repetition frequency.



### 8. Appendix

#### Effect of OSA sensitivity setting

The OSA sensitivity setting has a significant effect on the quality and duration of pulsed light measurements. In time average mode, the sensitivity setting limits the measurable pulse repetition frequency. On the other hand, in external trigger mode and peak hold mode, the minimum pulse width is limited. These are related to the response speed of the OSA detection circuit.

The OSA detection circuit is selected by the measurement sensitivity setting. Each circuit corresponds to each sensitivity setting and provides a specific gain and response speed. The gain and response speed are inversely related, and increasing the gain slows down the response speed. Therefore, at high sensitivity settings such as HIGH1, while the gain of the detection circuit increases, the reaction speed becomes slower and the wavelength sweep becomes slower. Conversely, with low sensitivity settings such as NORM-HOLD and NORMAL, the gain of the detection circuit is reduced, while the reaction speed is faster and the wavelength sweep is faster.

In time averaging mode, if the OSA responds too quickly compared to the pulse repetition frequency, there will be a clear on / off distortion in the recorded spectrum. Therefore, it is very important to select the appropriate measurement sensitivity for the pulse repetition frequency.

In external trigger mode and peak hold mode, peaks of short pulse width can be measured correctly at low sensitivity settings where the gain of detection circuit is low and the response is fast. However, at high sensitivity settings, the gain of the detection circuit is high and the response is slow, so short pulse width peaks cannot be measured correctly. Therefore, it is very important to select the appropriate measurement sensitivity for the pulse width to be measured.



#### OSA detection signal

- Laser output pulse signal
- Detection circuit output signal (low sensitivity)

In the low sensitivity setting, P ( $\lambda$ ) is high because the response of the detection circuit is fast. The peak of the pulse is captured.

• Detection circuit output signal (high sensitivity)

In the high sensitivity setting, P ( $\lambda$ ) is low because the response of the detection circuit is slow. The peak of the pulse cannot be captured.



#### Operation in peak hold mode



At each wavelength, the pulse detector signal is recorded at specific time intervals. The maximum signal acquired at the time interval is used as the power value at a specific wavelength.



## **Revision note**

Rev #	Descriptions
1	<ul> <li>Section 2 (Time Average Mode) Corrected the frequency for AQ6360 in the table of estimated minimum repetition frequency. (Changed from 2 Hz to 2 kHz)</li> <li>Section 7 (Examples) Changed the column title from average power to measured power and corrected the power calculation for all tables.</li> </ul>