

Specifications

AC Withstanding	
Output-Voltage Range	0.050kV~5.000kV
Output-Voltage Resolution	1V
Output-Voltage Accuracy	± (1% of setting + 5V) [no load]
Maximum Rated Load	150 VA (5kV/30mA)
Maximum Rated Current	30mA 0.001mA ~ 10mA (0.05kV ≤ V ≤ 0.5kV) 0.001mA ~ 30mA (0.5kV < V ≤ 5kV)
Output-Voltage Waveform	Sine wave
Voltage Regulation	± (1% + 5V) [maximum rated load → no load]
Output-Voltage Frequency	50 Hz / 60 Hz selectable
Voltmeter Accuracy	± (1% of reading + 5V)
Current Measurement Range	0.001mA ~ 30.00mA
Current Best Resolution	1µA (0.001mA ~ 9.999mA) 10µA (10.00mA ~ 30.00mA)
Current Measurement Accuracy	±(1.5% of reading + 30µA)
Current Offset	80µA maximum
ARC Detect	Yes
RAMP TIME (Rise Time)	0.1s~999.9s
FALL Time	OFF~999.9s
WAIT Time	OFF~999.9s
TIMER (Test Time)	CONT ² , 0.3s~999.9s
TIMER Accuracy	± (100ppm + 20ms)
GND	ON/OFF
DC Withstanding	
Output-Voltage Range	0.050kV~6.000kV
Output-Voltage Resolution	1V
Output-Voltage Accuracy	± (1% of setting + 5V) [no load]
Maximum Rated Load	50W (5kV/10mA)
Maximum Rated Current	10mA 0.001mA ~ 2mA (0.05kV ≤ V ≤ 0.5kV) 0.001mA ~ 10mA (0.5kV < V ≤ 6kV)
Voltmeter Accuracy	± (1% of reading + 5V)
Voltage Regulation	± (1% + 5V) [maximum rated load → no load]
Current Measurement Range	0.001mA ~ 10.00mA
Current Best Resolution	0.1µA (0.1µA ~ 999.9µA) 1µA (1µA ~ 9.999mA) 10µA (10.00mA)
Current Measurement Accuracy	±(1% of reading + 1µA) when I Reading < 1mA ±(1% of reading + 10µA) when I Reading ≥ 1mA
Current Offset	5µA maximum
ARC Detect	Yes
RAMP TIME (Rise Time)	0.1s~999.9s
FALL Time	OFF~999.9s
WAIT Time	OFF~999.9s
TIMER (Test Time)	CONT ² , 0.3s~999.9s
TIMER Accuracy	± (100ppm + 20ms)
GND	ON/OFF
Insulation Resistance	
Output Voltage	50V~1000V dc
Output-Voltage Resolution	1V
Output-Voltage Accuracy	± (1% of setting + 5V) [no load]
Resistance Measurement	0.1MΩ~ 10GΩ

Test Voltage	Measurement Range / Accuracy
50V \leq V < 500V	0.1MΩ~1MΩ : $\pm(5\% \text{ of reading} + 3 \text{ count})$ 1 MΩ~50MΩ : $\pm(5\% \text{ of reading} + 1 \text{ count})$ 51MΩ~2GΩ : $\pm(10\% \text{ of reading} + 1 \text{ count})$
500V \leq V \leq 1000V	0.1MΩ~1MΩ : $\pm(5\% \text{ of reading} + 3 \text{ count})$ 1 MΩ~500MΩ : $\pm(5\% \text{ of reading} + 1 \text{ count})$ 501MΩ~10GΩ : $\pm(10\% \text{ of reading} + 1 \text{ count})$
Voltage Regulation	$\pm(1\% + 5V)$ [maximum rated load \rightarrow no load]
Voltmeter Accuracy	$\pm(1\% \text{ of reading} + 5V)$
Short-Circuit Current	10mA max.
Output Impedance	2kΩ
RAMP TIME (Rise Time)	0.1s~999.9s
FALL Time	OFF~999.9s
WAIT TIME	OFF~999.9s
TIMER (Test Time)*	0.3s~999.9s
TIMER Accuracy	$\pm(100\text{ppm} + 20\text{ms})$
GND	ON/OFF
Continuity Test	
Output-Current	100mA dc
Ohmmeter Measurement Accuracy	$1\Omega \pm 0.2\Omega$, ON/OFF
Interface	
Signal I/O	Standard
RS-232C	Standard
USB (device)	Standard
USB (host)	Standard (for parameter / LCD hardcopy)
Rear Output	Scanner
DISPLAY	
	4.3" color LCD
POWER SOURCE	
	AC 100V~240V $\pm 10\%$, 50Hz/60Hz
DIMENSION & WEIGHT	
	320(W) x 120(H) x 435(D) mm; Approx. 11kg